

# SPEC PTDaemon Minor Update Certificate



ISO/IEC 21836:2020

Acceptance Scope PTDaemon 1.11.3

Compliance Evaluator Mike Petrich (IBM)  
ISG Server Efficiency Webmaster

Controller OS RHEL 8.10  
Power Analyzer Model Yokogawa WT310e  
Temperature Sensor Model Digi International Watchport /T  
Base PTDaemon version 1.8.1  
New PTDaemon version 1.11.3

Controller OS MS Windows 11 Enterprise 10.0.26100 Build 26100  
Power Analyzer Model Hioki PW3337  
Temperature Sensor Model Digi Interational Watchport /T  
Base PTDaemon version 1.8.1  
New PTDaemon version 1.11.3

| ISO/IEC 21836 Requirement        | System: Configuration Summary                                                                              | Pass / Fail |
|----------------------------------|------------------------------------------------------------------------------------------------------------|-------------|
| 9.3.4 PTDaemon Sum of Avg. Watts | Yokogawa WT310e (RHEL 8.10), ptdaemon 1.11.3 vs. 1.8.1                                                     | PASS        |
| 9.3.4 PTDaemon Avg. Temp. Diff.  | Digi International Watchport /T (RHEL 8.10) ptdaemon 1.11.3 vs. 1.8.1                                      | PASS        |
| 9.3.4 PTDaemon Sum of Avg. Watts | Hioki PW3337 (MS Windows 11 Enterprise 10.0.26100 Build 26100), ptdaemon 1.11.3 vs. 1.8.1                  | PASS        |
| 9.3.4 PTDaemon Avg. Temp. Diff.  | Digi Interational Watchport /T (MS Windows 11 Enterprise 10.0.26100 Build 26100) ptdaemon 1.11.3 vs. 1.8.1 | PASS        |

The data behind this certificate can be request requested via [sertsupport@spec.org](mailto:sertsupport@spec.org). Copyright © 2006-2025 Standard Performance Evaluation Corporation (SPEC). All rights reserved, see <https://www.spec.org>. SPEC, the SPEC logo, and the names SERT, PTDaemon, SPECpower are registered trademarks of SPEC. Additional product and service names mentioned herein may be the trademarks of their respective owners.